

Patient Monitor

M Series 10/M Series 12

Patient Monitor

10 inch



12 inch



M Series 10/M Series 12

Size and Weight

Size	M Series 12: 198mm X 320mm X 262mm M Series 10: 193mm X 288mm X 236mm
Weight	M Series 12: < 4kg M Series 10: < 3kg

Power

Standard According to	IEC 60601-1 and IEC 60601-1-2
Input voltage	AC (100-240) V(±10%)
Frequency	50Hz/60Hz
Input power	100VA

Display

Type	Color TFT LCD
Size(diagonal)	12.1" / 10.4" (M12 / M10)
Resolution	M Series 12 1280×800 pixels M Series 10: 1024×600 pixels

Recorder(M Series 12 Option)

Type	Thermal dot array (BTR50S)
Paper width	50 mm ±1mm
Recording speed	12.5 mm/s, 25 mm/s, 50 mm/s
Recording waveform	Maximum 3 tracks

Battery

Type	Rechargeable Li-ion battery 11.1V 2.5Ah
Operating time	>240 minutes (2.5Ah)
(1 new and fully charged battery at 25°C temperature, connecting SpO2 sensor & NIBP work on AUTO mode for 30 minutes interval)	
Charge time	<8 hours(2.5Ah)

Data Storage

Alarm event	3000 groups and associated waveform
Trend	1800h, minimum resolution is 10min 180h, minimum resolution is 1min 6h, minimum resolution is 5s
ARR event	3000 groups and associated waveform
NIBP	2400 groups
Holographic waveform	72 hours

Interfacing & I/O devices

Shortcut Keys	NIBP Start/Stop, alarm reset, alarm pause, Freeze
Control Knob	1
Keyboard & Mouse	Support
Barcode Scanner	Support 1D barcode (USB connector)
Wired network	1 standard RJ45 interfaces
Wifi (option)	Protocol: IEEE802.11a/b/g/n
Wifi frequency	Dual Band: 2.4G/5G
USB socket	2 sockets
Video output	1 VGA (option)
Multifunctional port	nurse call / defibrillation sync. / analog output

ECG

Lead	3 lead: I, II, III 5 lead: I, II, III, aVR, aVL, aVF, Vx 6-lead: I, II, III, aVR, aVL, aVF, Va, Vb Auto: identify leads automatically
Lead standard	AHA, IEC
Gain	Auto, 2.5 mm/Mv (×0.25), 5 mm/mV (×0.5), 10 mm/mV (×1), 20 mm/mV (×2), 40 mm/mV (×4)
CMRR	Monitor / Operation mode ≥ 110 dB Diagnostic mode ≥ 100 dB
Bandwidth (-3dB)	Monitor mode: 0.5 Hz to 40 Hz Operation mode: 1 Hz to 25Hz Diagnostic mode: 0.05Hz~150Hz ST mode: 0.05Hz~40Hz
Input impedance	≥ 5.0 MΩ
Input signal range	-10.0mV~+10.0mV
Electrode offset potential	± 500 Mv d.c.
System noise	≤ 30 μVpp (RTI)
Recovery time after defibrillation:	waveform recover to baseline in 10s
Sweep speed	6.25mm/s, 12.5 mm/s, 25 mm/s, 50mm/s.
ST segment	
Measurement range	-2.0 mV to +2.0 mV
Accuracy	-0.8 mV to +0.8 mV: ±0.02 mV or ±10% (whichever is greater)
Resolution	0.01mV
Heart Rate	
Measurement range	Adult 10 bpm to 300 bpm Pediatric & Neonatal 10 bpm to 350 bpm
Resolution	1 bpm
Accuracy	±1% or ±1 bpm, whichever is greater
Arrhythmia analysis	
27 Kinds	Asystole, Vent Fib/Tach, V-Tach, Vent Brady, Extreme Tachy,Extreme Brady,R on T,Tachy,Brady,Nonsustained V-Tach, Vent Rhythm, PNC, PNP, Pause, Pauses/min High, Run PVCs, Couplet, Bigeminy, Trigeminy, Frequent PVCs, PVC, Missed Beat, A-Fib, A-Fib End, ECG Noise, Irregular Rhythm, Irregular RhythmEnd.
Respiration	
Lead	Selected from: I (RA-LA) or II (RA-LL)
Measurement range	0 rpm to 150 rpm
Resolution	1 rpm
Accuracy	±2 rpm or ±2% , whichever is the greater
Delay of apnea alarm	Adjustable delay time: 10s ~ 60s

NIBP

Measurement way	Automatic oscillometry
Measurement mode	Manual , Auto, STAT, Sequence
Intervals for Auto measurement:	1/2/2.5/3/5/10/15/20/30min, 1/1.5/2/3/4/8h
STAT mode cycle time	5 minutes.
Sequence mode	Up to 5 group, and each group individually sets the interval and number of periodic measurement.

Systolic range	Adult	30 to 270 mmHg
	Pediatric	30 to 235 mmHg
	Neonatal	30 to 135 mmHg
Diastolic range	Adult	10 to 220 mmHg
	Pediatric	10 to 220 mmHg
	Neonatal	10 to 110 mmHg
Mean range	Adult	20 to 235 mmHg
	Pediatric	20 to 235 mmHg
	Neonatal	20 to 125 mmHg
Pressure accuracy	Static:	±3 mmHg (±0.4kPa)
	Clinic:	mean error ±5 mmHg
	Standard deviation:	≤8 mmHg
PR range	40 bpm to 240 bpm	
PR accuracy	±3bpm or ±3%, whichever is greater	
Measurement time	20s to 45s (typical value)	
Software overpressure protection	Adult	(297±3) mmHg
	Pediatric	(252±3) mmHg
	Neonatal	(147±3) mmHg

IBP (option for M Series 12 only)

Sensitivity of transducer	5uV/V/ mmHg, ±2%	
Impedance of transducer	300Ω to 3000Ω	
Measurement range	-50 mmHg to +360 mmHg	
Measurement accuracy	±2 mmHg or ±2% of the reading, whichever is the greater (exclusive of transducer)	
Resolution	1 mmHg	
Unit	mmHg, kPa, cmH2O	
Transducer sites	ART/CVP/ICP/PA/Ao/UAP/BAP/FAP//LAP/RAP/UVP LV/PAWP, additionally, P1 & P2 are arbitrary sites	
PPV		
Measurement range	0~50%	
Resolution	1.00%	
Measurement range	30 bpm to 300 bpm	
Resolution Accuracy	1bpm	
	±1% or ±1bpm whichever is greater	
Software overpressure protection	Adult	(297±3) mmHg
	Pediatric	(252±3) mmHg
	Neonatal	(147±3) mmHg

Temperature (Dual-Temp only for M Series 12)

Parameter	T1,T2,TD
Probe	YSI400 series probe (2252 Ω @25°C)
Measurement range	0.0°C to 50.0°C(32°F to 122 °F)
Accuracy	±0.1°C or ±1°F (exclusive of probe)
Resolution	0.1°C or 1°F
Unit	°C or °F

MicroFlow CO2 (option for M Series 12 only)

Measurement range	0% to 25% (0 mmHg to 190 mmHg)
Unit	0.1% or 1mmHg
Unit	%, mmHg, kPa
Accuracy	± (0.43% + 8% of reading)
Preheating time	<10s (Report concentration and achieve highest accuracy)
Rise time	<3s (including delay time and rise time)
Sample Flow Rate	50±10mL/min
awRR range	0 rpm to 150 rpm
awRR accuracy	±1 rpm

BLT SpO2

Measurement range	0% ~ 100%
Accuracy(clinical)	70% ~ 100% ≤3% (SpO2 probe included)
	0% ~ 69% unspecified

PR

Measurement range	25 bpm to 300 bpm
Resolution	1bpm
Accuracy	± 3bpm

PI

Measurement range	0.05~20.00%
Resolution	0.01%
Accuracy	±0.1% or ±10% of reading, whichever is greater

RESP (from pleth)

Measurement range	0 rpm ~90 rpm
Resolution	1 rpm
Accuracy	± 2rpm

Mainstream CO2 (option for M Series 12 only)

Measurement range	0% to 25% (0 mmHg to 190 mmHg)
Resolution	0.1% or 1mmHg
Preheating time	<10s
Rise time	<90ms
Unit	%, mmHg, kPa
Accuracy	± (0.43% + 8% of reading)
awRR range	0 rpm to 150 rpm
awRR accuracy	±1 rpm

Standard configuration :

3/5/6 lead ECG, HR, SpO2, PI, RESP(from pleth), NIBP, Temp, Dual-Temp(M Series 12), Rechargeable Li-ion battery (2.5Ah).

Option :

M Series 12: 2-IBP, Mainstream/Microflow EtCO2, Touch Screen, Thermal Printer.

M Series 10/M Series 12: Rolling stand,Wall mount.



*Specifications subject to change without prior notice.

AKAIMED LIMITED LTD.
706, 7/F, South Seas Centre, Tower 2, 75 Mody Road,
TsimSha Tsui, Kowloon, Hong Kong
Web: www.akaimed.com
Email: international@akaimed.com

AKAI Med